



Tritium Laboratory  
4 September 2024

SWAB REPORT # 1127

SWAB DATE: 29 August 2025

*R/V Pelican*

---

James D. Happell

Distribution:  
SWAB Committee  
Joseph Malbrough

## COMMENTS TO SWAB REPORTS

15 December 2021

The LSC is now a Quantulus GCT 6220, with the SWAB counting assay having background cpm of 0.3 & 1.2 for  $^3\text{H}$  &  $^{14}\text{C}$ . This replaces an LSC with background cpm of 1.6 & 5.5 for  $^3\text{H}$  &  $^{14}\text{C}$ .

All samples are counted for 60 minutes, the instrument background is subtracted, and activities are reported in  $\text{dpm/m}^2$ . Bucket blank activities are not subtracted. Counting errors (2 standard deviations) are also reported in  $\text{dpm/m}^2$ . An error larger than the activity indicates that the activity is not significantly different from zero. All activities significantly above background will be in **bold**.

### Criteria for SWAB Results

Category	$^3\text{H}$ ( $\text{dpm/m}^2$ )	$^{14}\text{C}$ ( $\text{dpm m}^2$ )	Recommendations
A	<500	<50	No action
B*	500-10,000	50-10,000	Needs cleaning before any natural tracer work. Decks in radiation vans with activities above $1000 \text{ dpm/m}^2$ should be cleaned.
C**	10,000-100,000	10,000-50,000	Must be cleaned before any use.
D***	>100,000	>50,000	May be a health hazard. Notify local radiation safety official.

Note:  $^{14}\text{C}$  and  $^{35}\text{S}$  have peak energies of 156 and 167 KeV, respectively; thus  $^{35}\text{S}$  will be registered as  $^{14}\text{C}$  by our counting techniques. Categories A, B and C are not a health hazard.

### Recommended Cleaning Procedure

Wearing ordinary household rubber gloves:

$^3\text{H}$ : Wash and scrub with radioactive cleanup detergent such as COUNT-OFF (50 ml COUNT-OFF to 4 liters of water), using sponges to distribute solution and reabsorb it.

$^{14}\text{C}$ : Wash with 1% sulfuric or 2% hydrochloric (muriatic) acid with good ventilation (will dissolve carbonates, releasing  $^{14}\text{CO}_2$ ). Follow up with wash as if for  $^3\text{H}$ .

### Disposal of Cleaning Materials (gloves, sponges, etc)

Categories A & B dispose as ordinary garbage, C & D contact your institution's radiation safety office.

Note: If category C or D is encountered, we try to notify the institution promptly by phone or email.

# REPORT FOR SWAB # 1127

LOCATION: Cocodrie, LA

VESSEL: *R/V Pelican*

DATE: 29 August 2025

TECHNICIAN: Joseph Gonzalez

Sample #	Sample Identification	<sup>3</sup> H dpm/m <sup>2</sup>		<sup>14</sup> C dpm/m <sup>2</sup>	
		activity	error	activity	error
1	1st Vial Bkgnd	0	± 0	0	± 0
2	Initial bucket blank	29	± 20	-24	± 59
	<u>Dry Lab (Figure 1)</u>				
3	Forward benchtop	53	± 22	-13	± 32
4	Port benchtop	49	± 20	-5	± 17
5	Forward benchtop next to fume hood	26	± 18	-14	± 34
6	Deck in front of door	50	± 27	-6	± 21
7	Fume hood	66	± 25	-17	± 41
	<u>Main Deck (Figure 1)</u>				
8	Deck outside computer room	57	± 23	-11	± 27
9	Deck between galley and mess	36	± 20	-10	± 25
10	Deck at top of forward stairs	62	± 25	-13	± 31
	<u>Wet Lab (Figure 1)</u>				
11	Deck in front of aft freezer	66	± 28	1	± 5
12	Inside forward freezer top	23	± 16	-6	± 19
13	Inside forward refrigerator bottom	86	± 22	4	± 7
14	Benchtop next to port door	39	± 19	-6	± 21
15	Benchtop next to forward sink	24	± 15	-1	± 89
16	Deck in front of refrigerator	47	± 31	-14	± 35
17	Benchtop across from refrigerator	5	± 8	0	± 10
18	Benchtop across from port sink	27	± 16	-7	± 23
19	Sink area of bottle lab	20	± 21	-20	± 48
20	Aft deck of bottle lab	44	± 18	2	± 8
21	Benchtop forward of port sink	7	± 43	-13	± 32
22	Deck inside port entrance	57	± 34	-12	± 29
23	Deck in front of port sink	57	± 32	-7	± 23
24	Top of aft chest freezer	36	± 18	-3	± 11
25	Final bucket blank	52	± 20	4	± 9

### Comments

Please note that the error reported for each isotope is the two-standard deviation counting error. Reports may now contain values less than zero. Decay counting background samples will be distributed about the background vial, which means that negative values are possible. In the past we rounded the negative values to zero. Values are only significantly above background when they are positive and larger than the error. Please note that we are now using a Quantulus 6220 LSC which counts very near natural background. While the cleanup standards have not changed all values above background will now be in bold. All areas tested on the ship were free from isotope contamination that requires cleaning.

# *R/V Pelican*

Figure 1  
SWAB #1127  
29 August 2025

