



7 October 2026

SWAB REPORT #1147

SWAB DATE: 5 October 2026

University of South Florida

James Happell

Digitally signed by James Happell  
Date: 2026.10.07 17:18:03 -04'00'

---

James D. Happell  
Associate Research Professor

Distribution:  
UNOLS office  
Justice Saxby

## COMMENTS TO SWAB REPORTS

15 December 2021

The LSC is now a Quantulus GCT 6220, with the SWAB counting assay having background cpm of 0.3 & 1.2 for  $^3\text{H}$  &  $^{14}\text{C}$ . This replaces an LSC with background cpm of 1.6 & 5.5 for  $^3\text{H}$  &  $^{14}\text{C}$ .

All samples are counted for 60 minutes, the instrument background is subtracted, and activities are reported in  $\text{dpm/m}^2$ . Bucket blank activities are not subtracted. Counting errors (2 standard deviations) are also reported in  $\text{dpm/m}^2$ . An error larger than the activity indicates that the activity is not significantly different from zero. All activities significantly above background will be in **bold**.

### Criteria for SWAB Results

| Category | $^3\text{H}$ ( $\text{dpm/m}^2$ ) | $^{14}\text{C}$ ( $\text{dpm m}^2$ ) | Recommendations                                                                                                                        |
|----------|-----------------------------------|--------------------------------------|----------------------------------------------------------------------------------------------------------------------------------------|
| A        | <500                              | <50                                  | No action                                                                                                                              |
| B*       | 500-10,000                        | 50-10,000                            | Needs cleaning before any natural tracer work. Decks in radiation vans with activities above $1000 \text{ dpm/m}^2$ should be cleaned. |
| C**      | 10,000-100,000                    | 10,000-50,000                        | Must be cleaned before any use.                                                                                                        |
| D***     | >100,000                          | >50,000                              | May be a health hazard. Notify local radiation safety official.                                                                        |

Note:  $^{14}\text{C}$  and  $^{35}\text{S}$  have peak energies of 156 and 167 KeV, respectively; thus  $^{35}\text{S}$  will be registered as  $^{14}\text{C}$  by our counting techniques. Categories A, B and C are not a health hazard.

### Recommended Cleaning Procedure

Wearing ordinary household rubber gloves:

$^3\text{H}$ : Wash and scrub with radioactive cleanup detergent such as COUNT-OFF (50 ml COUNT-OFF to 4 liters of water), using sponges to distribute solution and reabsorb it.

$^{14}\text{C}$ : Wash with 1% sulfuric or 2% hydrochloric (muriatic) acid with good ventilation (will dissolve carbonates, releasing  $^{14}\text{CO}_2$ ). Follow up with wash as if for  $^3\text{H}$ .

### Disposal of Cleaning Materials (gloves, sponges, etc)

Categories A & B dispose as ordinary garbage, C & D contact your institution's radiation safety office.

Note: If category C or D is encountered, we try to notify the insitution promptly by phone or email.

## REPORT FOR SWAB #1147

LOCATION: St. Petersburg, FL

DATE: 6 October 2026

VESSEL/LAB: University of South Florida

TECHNICIAN: Brad Rosenheim

| Sample # | Sample Identification          | <sup>3</sup> H dpm/m <sup>2</sup> |       | <sup>14</sup> C dpm/m <sup>2</sup> |             |
|----------|--------------------------------|-----------------------------------|-------|------------------------------------|-------------|
|          |                                | activity                          | error | activity                           | error       |
| 1        | 1st Vial Bkgnd                 | 0                                 | ± 0   | 0                                  | ± 0         |
| 2        | Initial bucket blank           | 31                                | ± 30  | 4                                  | ± 11        |
| 3        | Workbench near window 4010F    | -36                               | ± 56  | 2                                  | ± 49        |
| 4        | Workbenches with shelves 4010F | -19                               | ± 29  | -15                                | ± 25        |
| 5        | 4019C casework surface         | -13                               | ± 21  | 10                                 | ± 16        |
| 6        | 4019B casework surface         | -19                               | ± 30  | 4                                  | ± 18        |
| 7        | 4029 hoods and casework        | -28                               | ± 45  | -2                                 | ± 36        |
| 8        | 4019A casework surface         | 7                                 | ± 13  | <b>18</b>                          | ± <b>15</b> |
| 9        | 4018B cold room floor          | -25                               | ± 40  | -4                                 | ± 11        |
| 10       | 4010F floor                    | 29                                | ± 45  | -20                                | ± 33        |

## COMMENTS

Please note that the error reported for each isotope is the two-standard deviation counting error. Reports may now contain values less than zero. Decay counting background samples will be distributed about the background vial, which means that negative values are possible. In the past we rounded the negative values to zero. Values are only significantly above background when they are positive and larger than the error. Please note that we are now using a Quantulus 6220 LSC which counts very near natural background. While the cleanup standards have not changed all values above background will now be in bold. All areas tested were free from isotope contamination that requires cleaning.