

UNIVERSITY OF MIAMI  
ROSENSTIEL  
SCHOOL of MARINE &  
ATMOSPHERIC SCIENCE



Tritium Laboratory  
14 January 2014

---

Tritium Laboratory      Ph: 305-421-4100  
4600 Rickenbacker Causeway      Fax: 305-421-4112  
Miami, Florida 33149-1031      E-mail: Tritium@rsmas.miami.edu

SWAB REPORT # 714

SWAB DATE: 22 December 2013

*R/V Thomas Thompson*

---

Dr. James D. Happell  
Associate Research Professor

Distribution:  
SWAB Committee

## COMMENTS TO SWAB REPORTS

23 November 2010

Typical LSC instrument background values for  $^3\text{H}$  and  $^{14}\text{C}$  are 2 and 5 cpm, respectively. The LSC is a Tricarb 2910 TR with the low level counting option.

All samples are counted for 60 minutes, the instrument background is subtracted, and activities are reported in  $\text{dpm}/\text{m}^2$ . Bucket blank activities are not subtracted. Counting errors (2 standard deviations) are also reported in  $\text{dpm}/\text{m}^2$ . An error larger than the activity indicates that the activity is not significantly different from zero.

### Criteria for SWAB Results

Category	$^3\text{H}$ ( $\text{dpm}/\text{m}^2$ )	$^{14}\text{C}$ ( $\text{dpm m}^2$ )	Recommendations
A	<500	<50	No action
B*	500-10,000	50-10,000	Needs cleaning before any natural tracer work. Decks in radiation vans with activities above 1000 $\text{dpm}/\text{m}^2$ should be cleaned.
C**	10,000-100,000	10,000-50,000	Must be cleaned before any use.
D***	>100,000	>50,000	May be a health hazard. Notify local radiation safety official.

Note:  $^{14}\text{C}$  and  $^{35}\text{S}$  have peak energies of 156 and 167 KeV, respectively; thus  $^{35}\text{S}$  will be registered as  $^{14}\text{C}$  by our counting techniques. Categories A, B and C are not a health hazard.

### Recommended Cleaning Procedure

Wearing ordinary household rubber gloves:

$^3\text{H}$ : Wash and scrub with radioactive cleanup detergent such as COUNT-OFF (50 ml COUNT-OFF to 4 liters of water), using sponges to distribute solution and reabsorb it.

$^{14}\text{C}$ : Wash with 1% sulfuric or 2% hydrochloric (muriatic) acid with good ventilation (will dissolve carbonates, releasing  $^{14}\text{CO}_2$ ). Follow up with wash as if for  $^3\text{H}$ .

### Disposal of Cleaning Materials (gloves, sponges, etc)

Categories A & B dispose as ordinary garbage, C & D dispose in radiation waste system.

Note: If category C or D is encountered, we try to notify the insitution promptly by phone or email.

REPORT FOR SWAB # 714

LOCATION: Tahiti  
VESSEL: R/V *Thomas Thompson*

DATE: 22 December 2013  
TECHNICIAN: Jim Happell

Sample #	Sample Identification	<sup>3</sup> H dpm/m <sup>2</sup>			<sup>14</sup> C dpm/m <sup>2</sup>		
		activity	±	error	activity	±	error
1	1st Vial Bkgnd	0	±	0	0	±	0
2	Initial bucket blank- Dawn	34	±	51	0	±	0
3	Hydro Lab aft bench	27	±	32	31	±	34
4	Hydro Lab Aft sink area	55	±	42	18	±	30

**Comments**

Please note that the error reported for each isotope is the two-standard deviation counting error.  
No <sup>3</sup>H and <sup>14</sup>C contamination detected, no action required.